

Surat Municipal Institute of Medical Education and Research

Opp. Bombay Market, Umarwada, Surat – 395 010 (Guj.)-India.

Medical Superintendent-Extn.1600

SMIMER Hospital

Phone Nos. 2368040 to 43

FAX No.(0261) 2360306



SMIMER/Hosp/Out/ No.2605

Date: 11 / 08 /2025

To,
M/s _____

Sub: Quotation for repairing/replacement the Various parts of Multiparamonitor, ECG Machine & Ventilator of Pediatric , Medicine , Tb & Chest & Surgery department of SMIMER Hospital.

Dear Sir,

Quotations are invited for item as per attached list and as per following terms & conditions.

1. The rates should be F.O.R. destination inclusive of all taxes & charges i.e. GST etc.
 2. Rate should be quoted as per serial number of list.
 3. The rejected items shall have to be removed by the supplier at their own cost.
 4. Overwriting, erasures or correction in quotations shall not be considered.
 5. Quotations are to be sent in sealed cover as above.
 6. Quotations received after due date will not be considered.
 7. Quotations sent by post or courier services is preferable.
 8. Guarantee/warranty should be mentioned in quotation.
 9. Supplier has to submit quality certificate & samples if requires.
 10. Effective goods will not be accepted.
- GST Clause For Construction/ Erection/ Commissioning/Installation /Repairs/ Maintenance/Renovation/ Fabrication Of Structure Including Building(Means All Works Contract/Turn Key Projects/Supply Of Material/Goods)
 - GST (Goods & Service Tax) has come in existence from 1st July, 2017. Contractor/ Successful Bidder is bound to pay any amount of GST prescribed by the Govt. of India as per the Terms of contract agreed upon during the course of execution of this Contract.

- During the course of execution of contract, if there is any change in Rate of GST(Goods & Service Tax) by the Government, the same shall be reimbursed / recovered separately by SMC, subject to **the submission of original Receipt/Proof for the amounts actually remitted by** the successful Tenderer /Contractor to the Competent Authority along with a Certificate from Chartered Accountant of Contractor / Successful Bidder certifying that the amount of GST paid to the Government and the same shall be intimated / submitted/claimed within 15 (fifteen) Days from the date of payment. Remittance of GST within stipulated period shall be the sole responsibility of the Successful Bidder/Contractor, failing which, SMC may recover the amount due, from any other payable dues with SMC and decision of Municipal Commissioner shall be final and binding on the Contractor / Successful Bidder in this regard. Further, the non-payment of GST to the Government may lead to the termination of the contract and forfeiture of Security Deposit / Performance Guarantee Amount.
- If imposition of any other new Taxes /Duties /Levies / Cess or any other incidentals etc. or any increase in the existing Taxes / Duties / Levies /Cess or any other incidentals etc. (Excluding GST) are imposed during the course of the contract, the same shall be borne by the Contractor / Successful Bidder only, in no case SMC shall be liable for the same.
- **The contractor will submit the invoice to the SMC having GSTIN of SMC mentioned therein and the taxes shall be shown separately on the face of the invoice so as to claim as ITC by SMC.**

Envelope should bear on top: " Quotation for repairing/replacement the Various parts of Multiparamonitor, ECG Machine & Ventilator of Pediatric , Medicine , Tb & Chest & Surgery department of SMIMER Hospital."

Quotations are to be sent in sealed cover addressed to office of Medical Superintendent, SMIMER Hospital by post/courier so as to reach on Dt. 21 / 08 /2025. Quotation received thereafter will not be accepted.

Medical Superintendent
SMIMER Hospital

Price-Bid

Sr. No.	Item Name	Qty.	Department	Unit Rate (Rs.)	Total Amount (Rs.)
1.	Multipara Monitor Make:Contec Model: CMS 8000 Sr. No. 21030400047, 21030400114 , 21030400021, 21030400225 ,21030400133 • Replacement Of Compatible Neonate SpO2 Probe. <ul style="list-style-type: none">• Compatible reusable Neonate Spo2 probe y-type with silicon wrap.• Should be supplied with extension cable and y-type probe both.• The neonate probe must be compatible with contec spo2 technology.• Total Length=3meter approx.• Spo2 Accuracy 70 to 100 % ± 3%• Should be ISO/CE/BSI certified products only.• One year comprehensive warranty• The bidders are requested to visit the site to check the unit for confirming the SPO2 technology before bidding.• The SpO2 probe, provided it is not physically damaged, shall be covered under the warranty. No exceptions or excuses will be accepted in this regard..	05	Paediatric (NICU)		
2.	Multipara Monitor Make:Comen Model:Star 8000 Sr. No.F7180313007B,F7180313008B, F7181227012B , F7181227013B • Replacement Of Compatible Neonate SpO2 Probe <ul style="list-style-type: none">• Compatible reusable neonate Spo2 probe y-type soft rubber /silicon type with wrap.• Must be Compatible with the above mentioned units Spo2 Technology.• Length approx 3 mtr.• Spo2 Accuracy 70 to 100 % ± 3% accuracy.• Should be ISO/CE/BSI certified products only.• One year comprehensive warranty.• The bidders are requested to visit the site to check the unit for confirming the SPO2 technology before bidding.• The SpO2 probe, provided it is not physically damaged, shall be covered under the warranty: No exceptions or excuses will be accepted in this regard.	04			
3.	Neonatal Ventilator Make:Schiller India Model:Graphnet Adv. Sr.No.180940371081A1V, 18094038A1V <ul style="list-style-type: none">• Replacement Of Oxygen Sensor• General Requirements :<ul style="list-style-type: none">• The oxygen sensor must be fully compatible with the Graphnet Advance ventilator (Schiller India) without requiring any modification to the existing unit.• It must be suitable for continuous use in neonatal care settings, specifically in the NICU environment.• The sensor must meet clinical accuracy standards suitable for neonatal patients.• Technical Specifications:• Type: Galvanic or Electrochemical (OEM or	02			

	<p>OEM-equivalent with proven compatibility)</p> <ul style="list-style-type: none">• Range: 0-100% O2 measurement capability• Accuracy: ±2% of full scale or better• Output Signal: Analog signal compatible with the Graphnet Advance interface• Response Time: ≤ 15 seconds (T90)• Operating Temperature Range: 0°C to 50°C• Operating Humidity Range: 0% to 95% RH (non-condensing)• Lifespan: Minimum operational life of 12 months under normal use conditions• Compatibility & Compliance:• Must be 100% compatible with Graphnet Advance ventilators supplied by Schiller India without the need for adapters, firmware updates, or calibration modifications.• Must meet or exceed international standards for oxygen sensors used in critical care environments (e.g., ISO, CE marked).• Warranty and Support :• The oxygen sensor must be supplied with a one-year comprehensive warranty covering manufacturing defects, functional failures, and premature performance degradation.• The supplier must offer prompt replacement in case of sensor failure within the warranty period.• Packaging and Delivery:• The sensor should be supplied individually sealed in sterile, protective packaging.• Packaging should clearly display the manufacturing and expiration dates.				
4.	<p>Multipara Monitor Model:Draeger Vista 120s Make:- Draeger Sr. No. V9SKE 0043, SISLE 0017 , SISLE 0010</p> <p>➤ Replacement Of Compatible reusable Clip Type/ Soft Rubber Adult SpO2 Probe.</p> <ul style="list-style-type: none">• Compatible reusable Pediatric Spo2 probe soft rubber or clip type.• Must be Compatible with the above mentioned units Spo2 Technology.• Length approx 3 mtr.• Spo2 Accuracy 70 to 100 % ± 3% accuracy.• Should be ISO/CE/BIS/CDSCO certified products only.• One year comprehensive warranty.• The bidders are requested to visit the site to check the unit for confirming the SPO2 technology before bidding.• The SpO2 probe, provided it is not physically damaged, shall be covered under the warranty. No exceptions or excuses will be accepted in this regard.•	03	Paediatric (PICU)		
5.	<p>Multipara Monitor Model:Draeger Vista 120s Make:- Draeger Sr. No. SISLJ0044, SISLE0006</p> <p>➤ Compatible reusable Clip Type/ Soft Rubber Adult SpO2 Probe.</p> <ul style="list-style-type: none">• Compatible reusable Pediatric Spo2 probe soft rubber or clip type.• Must be Compatible with the above mentioned units Spo2 Technology.• Cable Materialp TPU jacket Material & Latex Free.	02	Medicine (E4A ICU-1 nos.) Surgery (2BICU - 1 nos.)		

	<ul style="list-style-type: none"> Length approx 3 mtr. Spo2 Accuracy 70 to 100 % \pm 3% accuracy. Should be ISO/CE/BIS/CDSCO certified products only. One year comprehensive warranty. The bidders are requested to visit the site to check the unit for confirming the SPO2 technology before bidding. The SpO2 probe, provided it is not physically damaged, shall be covered under the warranty. No exceptions or excuses will be accepted in this regard. Replacement of 5 lead ECG cable (adult) Sr.no. SISLJ0027 ,SISLJ0046 Compatible reusable 5 ECG limb lead grabber type wires with trunk cable. Should have grabber type limb wires. Should have detachable limb wires(preferable) so that each wire can be replaced as and when required. Leads should be color coded as per IEC All the lead wires must be encased in a lockable socket(if detachable) Must be defibrillator shock proof. Should be ISO/CE/BIS certified. The bidders are requested to kindly check the above machine on site before bidding. One year comprehensive warranty 	02	Medicine (E5B)		
6.	<p>Multipara Monitor Model: Vivid Vue 12 Make: BPL Medical Technologies Sr. No. RFMG1K1436</p> <ul style="list-style-type: none"> Replacement Of Compatible Neonate SpO2 Probe Compatible reusable Neonate Spo2 probe y-type with silicon wrap. Should be supplied with extension cable and y-type probe both. The neonate probe must be compatible with the above units spo2 technology. Total Length=3meter approx. Spo2 Accuracy 70 to 100 % \pm 3% Should be ISO/CE/BSI certified products only. One year comprehensive warranty The bidders are requested to visit the site to check the unit for confirming the SPO2 technology before bidding. The SpO2 probe, provided it is not physically damaged, shall be covered under the warranty. No exceptions or excuses will be accepted in this regard. 	01			
7.	<p>Multipara Monitor Model: Oxyview Make: BPL Medical Technologies Sr. No. LYTAIF 1691 , LYTAIF 1692, LYTAIF 1694</p> <ul style="list-style-type: none"> Replacement Of Compatible Neonate SpO2 Probe Compatible reusable Neonate Spo2 probe y-type with silicon wrap. Should be supplied with extension cable and y-type probe both. 	03	Pediatric (NICU)		

	<ul style="list-style-type: none"> • The neonate probe must be compatible with the above units spo2 technology. • Total Length=3meter approx. • Spo2 Accuracy 70 to 100 % ± 3% • Should be ISO/CE/BSI certified products only. • One year comprehensive warranty • The bidders are requested to visit the site to check the unit for confirming the SPO2 technology before bidding. • The SpO2 probe, provided it is not physically damaged, shall be covered under the warranty. No exceptions or excuses will be accepted in this regard. 				
8.	<p>Multipara Monitor Model:C-50Make:COMEN Sr.noKD18031400813,KD180314009B,KD180314013B</p> <ul style="list-style-type: none"> • Replacement of Compatible Reusable NIBP cuff (Adult) • Compatible reusable NIBP Cuff (adult) with male connector pin. • Nylon, TPU material • Size: 25-35cm • Reusable with bladder • Hose length-1 ft approx. • Should be ISO/CE/BIS certified products only. • 6 months comprehensive warranty. • Replacement of 5 lead ECG cable (adult) • Sr.no. KD180314008B , KD180314009B • Compatible reusable 5 lead ECG limb lead wires (ADULT) with trunk cable. • Should have grabber type limb wires. • Should have detachable limb wires (preferable) so that each wires can be • replaced as and when required. • Leads should be color coded as per IEC. • Should be ISO/CE/BIS certified products only. • One year comprehensive warranty. 	03	TB & Chest (RICU)		
9.	<p>3 Channel ECG Machine Make:Comen Model: CM300 Sr. No.30200119005</p> <ul style="list-style-type: none"> • Replacement of Compatible 10 Lead ECG Patient cable • Compatible reusable 10 lead banana pin type ECG patient cable. • The ECG cable should be labeled with standard electrode names for easy identification. • Should follow IEC (European Standard) color coding standards for ECG cable. • The 10 lead ECG cable should be provided with junction for replacing/repairing lead wires easily (Preferable) • Should be ISO/CE/BIS/FDA certified. • Warranty: - One Year comprehensive Warranty. If the ECG cable sheath starts to worn out during the warranty period the same must be covered in the warranty and it will not be considered as physical damage. • The bidders are requested to check the unit onsite for compatibility of the ECG cable. 	01	Surgery MSW (D4A)		

Note: Above mentioned Instrument should be physically checked in, Department Pediatric , Medicine , Tb & Chest & Surgery by the Quoting supplier prior to Quoting in SMIMER Hospital.

Medical Superintendent
SMIMER Hospital