

Surat Municipal Institute of Medical Education and Research

Opp. Bombay Market, Umarwada, Surat – 395 010 (Guj.)-India.

Medical Superintendent-Extn.1600

SMIMER Hospital

Phone Nos. 2368040 to 43

FAX No.(0261) 2360306



SMIMER/Hosp/Out/ No.2868

Date: 28 / 08 /2025

To,
M/s _____

Sub: Quotation for repairing/replacement the Various parts of Multiparamonitor & ECG Machine Medicine , Emergency & Anaesthesia department of SMIMER Hospital.

Dear Sir,

Quotations are invited for item as per attached list and as per following terms & conditions.

1. The rates should be F.O.R. destination inclusive of all taxes & charges i.e. GST etc.
 2. Rate should be quoted as per serial number of list.
 3. The rejected items shall have to be removed by the supplier at their own cost.
 4. Overwriting, erasures or correction in quotations shall not be considered.
 5. Quotations are to be sent in sealed cover as above.
 6. Quotations received after due date will not be considered.
 7. Quotations sent by post or courier services is preferable.
 8. Guarantee/warranty should be mentioned in quotation.
 9. Supplier has to submit quality certificate & samples if requires.
 10. Effective goods will not be accepted.
- GST Clause For Construction/ Erection/ Commissioning/Installation /Repairs/ Maintenance/Renovation/ Fabrication Of Structure Including Building(Means All Works Contract/Turn Key Projects/Supply Of Material/Goods)
 - GST (Goods & Service Tax) has come in existence from 1st July, 2017. Contractor/ Successful Bidder is bound to pay any amount of GST prescribed by the Govt. of India as per the Terms of contract agreed upon during the course of execution of this Contract.

- During the course of execution of contract, if there is any change in Rate of GST(Goods & Service Tax) by the Government, the same shall be reimbursed / recovered separately by SMC, subject to **the submission of original Receipt/Proof for the amounts actually remitted by the successful Tenderer /Contractor** to the Competent Authority along with a Certificate from Chartered Accountant of Contractor / Successful Bidder certifying that the amount of GST paid to the Government and the same shall be intimated / submitted/claimed within 15 (fifteen) Days from the date of payment. Remittance of GST within stipulated period shall be the sole responsibility of the Successful Bidder/Contractor, failing which, SMC may recover the amount due, from any other payable dues with SMC and decision of Municipal Commissioner shall be final and binding on the Contractor / Successful Bidder in this regard. Further, the non-payment of GST to the Government may lead to the termination of the contract and forfeiture of Security Deposit / Performance Guarantee Amount.
- If imposition of any other new Taxes /Duties /Levies / Cess or any other incidentals etc. or any increase in the existing Taxes / Duties / Levies /Cess or any other incidentals etc. (Excluding GST) are imposed during the course of the contract, the same shall be borne by the Contractor / Successful Bidder only, in no case SMC shall be liable for the same.
- **The contractor will submit the invoice to the SMC having GSTIN of SMC mentioned therein and the taxes shall be shown separately on the face of the invoice so as to claim as ITC by SMC.**

Envelope should bear on top: " Quotation for repairing/replacement the Various parts of Multiparamonitor & ECG Machine Medicine , Emergency & Anaesthesia department of SMIMER Hospital."

Quotations are to be sent in sealed cover addressed to office of Medical Superintendent, SMIMER Hospital by post/courier so as to reach on Dt. 08 / 09 /2025. Quotation received thereafter will not be accepted.

Medical Superintendent
SMIMER Hospital

Price-Bid

Sr. No.	Item Name	Qty.	Department	Unit Rate (Rs.)	Total Amount (Rs.)
1.	Multipara Monitor Make:Contec Model: CMS 8000 Sr. No. 21030400178, 21030400248, 2103040025,21012600173 <ul style="list-style-type: none"> • Replacement of Reusable NIBP cuff (Adult) <ul style="list-style-type: none"> ➤ Compatible reusable NIBP Cuff (adult) with male connector pin ➤ Nylon, TPU material ➤ Size:-25-35cm ➤ Reusable with bladder ➤ Hose Length -1 ft Approx. ➤ Should be ISO/CE/BIS certified product. 6 months Comprehensive warranty	04	Medicine		
2.	Multipara Monitor Make:Comen Model: Comen C-80 Sr.No.KE190102020B <ul style="list-style-type: none"> • Replacement Of Adult SpO2 Probe • Compatible reusable adult Spo2 probe soft rubber /clip type. • Must be Compatible Comen monitor Spo2 Technology. • Length approx 3 mtr. • Spo2 Accuracy 70 to 90 % ± 3% accuracy. • Cable material: TPU jacket Material & Latex free. • Should be ISO/CE/BSI certified products only. • One year comprehensive warranty. • The bidders are requested to visit the site to check the unit for confirming the SPO2 technology before bidding. • If physical cuts/ damages are there over the probe, the same must not be covered under warranty. 	01	Medicine (E4A ICU)		
3.	12 Channel ECG Machine Make :BPL Healthcare Pvt.Ltd. Model: Cardiart 9108 Sr.No.DUKT2835 <ul style="list-style-type: none"> • Replacement of Compitable 10 lead ECG Patient Cable <ul style="list-style-type: none"> ➤ Compitable reusable 10 lead banana pin type ECG patient cable. ➤ The ECG cable should be labeled with standard electrode names for easy identification. ➤ Should follow IEC (European Standard) color coding standard for ECG cable. ➤ The 10 lead ECG cable should be provide with junction for replacing/repairing lead wires easily & should not be with molded junction(Preferably) ➤ Should be ISO/CE/BIS/FDA certified. One Year Comprehensive warranty.	01	Emergency		
4.	3 Channel ECG Machine Make :Comen Model: CM-300 Sr.No.30200715038 <ul style="list-style-type: none"> • Replacement of Compitable 10 lead ECG Patient Cable 	01			

	<ul style="list-style-type: none"> ➤ Compatible reusable 10 lead banana pin type ECG patient cable. ➤ The ECG cable should be labeled with standard electrode names for easy identification. ➤ Should follow IEC (European Standard) color coding standard for ECG cable. ➤ The 10 lead ECG cable should be provide with junction for replacing/repairing lead wires easily & should not be with molded junction(Preferably) ➤ Should be ISO/CE/BIS/FDA certified. <p>One Year Comprehensive warranty.</p>				
5.	<p>3 Channel ECG Machine Make : BPL Healthcare Pvt.Ltd Model: 6208 View Plus Sr.No.DETD4C1665</p> <ul style="list-style-type: none"> • Replacement of ECG clamp Electrode(Color Coded) - 4 Electrodes=1 set <ul style="list-style-type: none"> ➤ Description: Reusable ECG clamp electrodes designed for limb placement (Right Arm, Left Arm, Right Leg, Left Leg). Color coded for easy identification and standard placement. ➤ Quantity : Set of 4 (RA - Red, LA - Yellow, RL - Black, LL - Green) ➤ Technical Specifications Type : Limb Clamp Type Material (Electrode): Stainless Steel / Nickel-Plated Copper Body Material : ABS Plastic (High Strength, Non-Conductive) Color Coding : Standard IEC color code: Red (RA), Yellow (LA), Green (LL), Black(RL) Connection Type : Standard 4mm banana plug or snap connector Electrode Area : Smooth, non-irritating surface Clamp Mechanism : Spring-loaded or pressure-fit clamp Reusability: Yes, autoclavable or disinfectable Compatibility: Universal - compatible with most ECG machines Certifications : CE or ISO 13485 or RoHS compliant ➤ Applications: Used for limb lead placement in ECG monitoring and diagnostic procedures. • Replacement of Chest Bulb Electrodes with Metal Electrodes (Silicone Type) 6 Electrodes Set = 1 Set <ul style="list-style-type: none"> ➤ Description : Precordial (chest) suction electrodes designed for use in ECG testing, equipped with soft silicone bulbs and metal contact points for accurate signal transmission. ➤ Quantity : Set of 6 chest electrodes (V1-V6) ➤ Technical Specifications 	<p>1 Set</p> <p>1 Set</p>	Anaesthesia (O.O.T)		

	<p>Type: Suction / Precordial Electrode</p> <p>Bulb Material : Medical-Grade Silicone (non-toxic, latex-free)</p> <p>Electrode Material: Stainless Steel / Nickel-plated Brass</p> <p>Electrode Size: Diameter: 2.5-3.0 cm (standard chest electrode size)</p> <p>Bulb Diameter: ~4.0-4.5 cm for optimal suction</p> <p>Connection Type: Standard 4mm banana plug or snap connector</p> <p>Color Coding: Optional: Color-coded or uniform gray/blue</p> <p>Suction Mechanism: Manual vacuum by squeezing the bulb</p> <p>Reusability : Yes, autoclavable or disinfectable</p> <p>Compatibility: Compatible with all standard 12-lead ECG systems</p> <p>Certifications: CE or ISO 13485 or RoHS compliant</p> <p>➤ Applications: Used for chest lead placement (VI to V6) in 12-lead ECG diagnostics.</p>				
--	---	--	--	--	--

Note: Above mentioned Instrument should be physically checked in, Department Medicine , Emergency & Anaesthesia by the Quoting supplier prior to Quoting in SMIMER Hospital.

Medical Superintendent
SMIMER Hospital