

Surat Municipal Institute of Medical Education and Research

Opp. Bombay Market, Umarwada, Surat – 395 010 (Guj.)-India.

Medical Superintendent-Extn.1600
SMIMER Hospital
Phone Nos. 2368040 to 43

FAX No.(0261) 2360306



SMIMER/Hosp/Out/ No.4626

Date: 04 / 10 /2025

To,
M/s _____

Sub: Quotation for repairing/replacement the Various parts of Various Instruments of Various department of SMIMER Hospital.

Dear Sir,

Quotations are invited for item as per attached list and as per following terms & conditions.

1. The rates should be F.O.R. destination inclusive of all taxes & charges i.e. GST etc.
 2. Rate should be quoted as per serial number of list.
 3. The rejected items shall have to be removed by the supplier at their own cost.
 4. Overwriting, erasures or correction in quotations shall not be considered.
 5. Quotations are to be sent in sealed cover as above.
 6. Quotations received after due date will not be considered.
 7. Quotations sent by post or courier services is preferable.
 8. Guarantee/warranty should be mentioned in quotation.
 9. Supplier has to submit quality certificate & samples if requires.
 10. Effective goods will not be accepted.
 11. Quoted rate should be valid for 1 year.
- GST Clause For Construction/ Erection/ Commissioning/Installation /Repairs/ Maintenance/Renovation/ Fabrication Of Structure Including Building(Means All Works Contract/Turn Key Projects/Supply Of Material/Goods)
 - GST (Goods & Service Tax) has come in existence from 1st July, 2017. Contractor/ Successful Bidder is bound to pay any amount of GST prescribed by the Govt. of India as per the Terms of contract agreed upon during the course of execution of this Contract.

- During the course of execution of contract, if there is any change in Rate of GST(Goods & Service Tax) by the Government, the same shall be reimbursed / recovered separately by SMC, subject to **the submission of original Receipt/Proof for the amounts actually remitted by** the successful Tenderer /Contractor to the Competent Authority along with a Certificate from Chartered Accountant of Contractor / Successful Bidder certifying that the amount of GST paid to the Government and the same shall be intimated / submitted/claimed within 15 (fifteen) Days from the date of payment. Remittance of GST within stipulated period shall be the sole responsibility of the Successful Bidder/Contractor, failing which, SMC may recover the amount due, from any other payable dues with SMC and decision of Municipal Commissioner shall be final and binding on the Contractor / Successful Bidder in this regard. Further, the non-payment of GST to the Government may lead to the termination of the contract and forfeiture of Security Deposit / Performance Guarantee Amount.
- If imposition of any other new Taxes /Duties /Levies / Cess or any other incidentals etc. or any increase in the existing Taxes / Duties / Levies /Cess or any other incidentals etc. (Excluding GST) are imposed during the course of the contract, the same shall be borne by the Contractor / Successful Bidder only, in no case SMC shall be liable for the same.
- **The contractor will submit the invoice to the SMC having GSTIN of SMC mentioned therein and the taxes shall be shown separately on the face of the invoice so as to claim as ITC by SMC.**

Each Quotation must be submit as per instruction mentioned below.

1. ENVELOPE NO.:1 (TECHNICAL BID)

In this envelope each Supplier has to submit necessary, Copy of Registration (If applicable), Experience Certificate, Detail of Parent Company or Approved agents of the parent co., Technical terms & conditions & other literature, details, User List of quoted products, Reference Letters from users of quoted products etc. duly sealed. On the envelope, title must be 'TECHNICAL BID', envelope which shall contents details of Supplier's Name, Address, Name of work and details of Quotation Notice.

2. ENVELOPE NO.2 (PRICE BID)

This Sealed cover should contain the supplier's "quoted rate" duly filled & signed and rubber stamped in the supplement given along with the quotation. The rate should be typed in figures as well as in words in the given price-bid format. The rates should be all-inclusive & meant for door-step delivery to SMC. There should be no hidden costs whatsoever. On the envelope title must be PRICE BID envelope which shall also contain details of supplier's Name, Address, Name of work and details of Quotation Notice.

3. ENVELOPE NO.3

Supplier has to submit above mentioned envelope No.1 & No.2 in this envelope duly sealed & this envelop should reach to OFFICE OF THE MEDICAL SUPERINTENDENT, SMIMER HOSPITAL, OPP. BOMBAY MARKET, UMARWADA, SURAT-395010 on or before due date with details of supplier's Name, Address, Name of work and details of Quotation Notice.

Envelope should bear on top: " Quotation for repairing/replacement the Various parts of Various Instruments of Various department of SMIMER Hospital ."

Quotations are to be sent in sealed cover addressed to office of Medical Superintendentshri, SMIMER Hospital by post/courier so as to reach on Dt. 10 / 10 /2025. Quotation received thereafter will not be accepted.

Medical Superintendent
SMIMER Hospital

Technical bid

DEPARTMENT OF RADIOLOGY			
ITEMWISE DETAILED SPECIFICATIONS :	Quoted / Not Quoted	Model	Manufacturer
(A) Required Specifications :			
Item No.1 :- Multipara Monitor Make: ComenModel: Comen C-50 ➤ Specification For Reusable NIBP cuff (Adult) <ul style="list-style-type: none">• Compatible reusable NIBP Cuff (adult) with male connector pin• Nylon, TPU material• Size:-25-35cm• Reusable with bladder• Hose Length -1 ft Approx.• Should be ISO/CE/BIS certified product. Warranty 6 months Comprehensive			
Item No.2 :- Multipara Monitor Make: ComenModel: Comen C-50 ➤ Specification For 5 lead ECG cable (Adult) <ul style="list-style-type: none">• Compatible reusable 5 lead ECG limb lead wires (adult) with trunk cable• Should have grabber type limb wires• Leads should be color coded as per IEC• Should be ISO/CE/BIS certified products only.• Bidders are requested to kindly visit the site before bidding to get a complete overview of the unit for ECG cable connections before bidding. One year Comprehensive warranty			
Item No.3 :- Multipara Monitor Make: Comen Model: Comen C-50 ➤ Specification For Reusable NIBP cuff (Adult) <ul style="list-style-type: none">• Compatible reusable NIBP Cuff (adult) with male connector pin• Nylon, TPU material• Size:-25-35cm• Reusable with bladder• Hose Length -1 ft Approx.• Should be ISO/CE/BIS certified product. Warranty 6 months Comprehensive warranty			

<p>Item No.4 :- Multipara Monitor Make: ContecModel: CMS 8000</p> <p>➤ Specification For Reusable NIBP cuff (Adult)</p> <ul style="list-style-type: none"> • Compatible reusable NIBP Cuff (adult) with male connector pin • Nylon, TPU material • Size:-25-35cm • Reusable with bladder • Hose Length -1 ft Approx. • Should be ISO/CE/BIS certified product. <p>Warranty 6 months Comprehensive warranty</p>			
<p>Item No.5 :- Multipara Monitor Make: ContecModel: CMS 8000</p> <p>➤ Specification For Compatible reusable Clip/Soft Rubber adult Spo2 probe</p> <ul style="list-style-type: none"> • The spo2 probe must be compatible with the above machines Spo2 technology • Cable length:- 3 meter approx • Cable material:- TPU jacket material & latex free • Spo2 Accuracy:- For 99 to 70% saturation ±2% accuracy • Should be ISO/CE/BIS certified product. <p>Warranty 1 year Comprehensive</p>			
<p>Item No.6 :- Multiparamonitor Model:- CMS 8000 Make:- Contec</p> <p>➤ Specification For 5 lead ECG cable (adult):-</p> <ul style="list-style-type: none"> • Compatible reusable 5 lead ECG limb lead wires (ADULT) with trunk cable. • Should have grabber type limb wires. • Leads should be color coded as per IEC. • Should be ISO/CE/BIS/ CDSCO certified products only. • One year comprehensive warranty. <p>The bidders are requested to kindly check the above said unit onsite before quoting.</p>			
<p>Item No.7 :- 3 Channel ECG Machine Make :comen Model: CM300</p> <p>➤ Specification For Compatible 10 lead ECG Patient Cable</p> <ul style="list-style-type: none"> • Compatible reusable 10 lead banana pin type ECG patient cable. • The ECG cable should be labeled with standard electrode names for easy identification. • Should follow IEC (European Standard) color coding standard for ECG cable. • The 10 lead ECG cable should be provide with junctionfor replacing/repairing lead wires easily & should not be with molded junction. • Should be ISO/CE/BIS/FDA certified. <p>One Year Comprehensive warranty</p>			

<p>Item No.8 :- 3 channel ECG Machine Make: BPL Healthcare Pvt.LtdModel:- Cardiart 6208 view plus</p> <p>➤ Specification For Compatible 10 lead ECG patient cable :-</p> <ul style="list-style-type: none"> • Compatible reusable 10 lead banana pin type ECG patient cable. • The ECG cable should be labeled with standard electrode names for easy identification. • should follow IEC(European Standard) Color coding standards for ECG cable. • The 10 lead ECG cable should be provided with junction for replacing/repairing lead wires easily & should not be with molded junction. • Should be ISO/CE/BIS certified product <p>Warranty 1 year Comprehensive</p>			
<p>Item No.9 :- Multipara Monitor Make: DragerModel: Vista 120S</p> <p>➤ Specification For NIBP cuff (Adult):-</p> <ul style="list-style-type: none"> • Compatible reusable NIBP cuff (Adult) with male connector pin. • Nylon,TPU material • Size: 25-35cm • Reusable with bladder • Hose length -1 ft approx. • Should be ISO/CE/BIS certified products only. <p>6 months comprehensive warranty.</p>			
<p>Item No.10 :- Multipara bedsidemonitor Make : DRAEGER Model: Vista 120S</p> <p>➤ Specification For Compatible reusable Clip Type/ Soft Rubber Adult SpO2 Probe</p> <ul style="list-style-type: none"> • Cable lenth:- 3 meter approx • Cable material:- TPU jacket material & latex free • Spo2 Accuracy:- For 99 to 70% saturation $\pm 3\%$ accuracy • Should be ISO/CE/BIS certified product. • Kindly visit the site for compelete over view of the machine/probe and spo2 technology <p>Warranty 1 year Comprehensive Warranty</p>			
<p>Item No.11 :- Multipara bedsidemonitor Make : DRAEGER Model: Vista 120S</p> <p>➤ Specification For Compatible 5 Lead ECG cable</p> <ul style="list-style-type: none"> • Compatible reusable 5 lead grabber type limb wires without trunk cable. • Should have detachable limb wires(preferable) so that each wire can be replaced as and when required. • All the lead wires must be encased in a lockable socket(if detachable) • Must be defibrilator shock proof. Also should be suitable while cautery machine is used in 			

<ul style="list-style-type: none"> operation theatre. Should follow color coding as per IEC. Should be ISO/CE/BIS/FDA certified. <p>One year comprehensive warranty</p>			
<p>Item No.12 :- Defibrillator Make: Nasan Model: Sajeevani 1006</p> <p>➤ Specification For Compatible Reusable NIBP cuff (Adult):-</p> <ul style="list-style-type: none"> Compatible reusable NIBP cuff (Adult) with male connector pin. Nylon,TPU material Size: 25-35cm Reusable with bladder Hose length -1 ft approx. Should be ISO/CE/BIS certified products only. <p>6 months comprehensive warranty.</p>			
<p>Item No.13:- Defibrillator Make: NasanModel:-Sanjeevani 1006</p> <p>➤ Specification For Rechrgeable battery</p> <ul style="list-style-type: none"> Compatible rechargeable lithium ion battery for above mention unit 11.1 Volt,5200 mAH The bidders are requested to kindly check the above machine on site before bidding. Should be ISO/CE/BIS certified product. <p>Warranty 1 year Comprehensive Warranty</p>			
<p>Item No.14 :- Defibrillator Make: Nasan Sr. no. 200653</p> <p>➤ Specification For Rechargeable Battery</p> <ul style="list-style-type: none"> Compatible rechargeable Li-ion battery for above mentioned unit. Battery details :- 11.1v,5200 mAh. Should be ISO/CE/BIS or equivalent certified Products Only. The bidders are requested to kindly check the above machine on site before bidding. <p>Warranty : 1 Year comprehensive</p>			
<p>Item No.15 :- Multipara Monitor Make: COMEN Model: C-80 Sr.</p> <p>➤ Specification For Compatible reusable Clip/Soft Rubber adult Spo2 probe</p> <ul style="list-style-type: none"> The spo2 probe must be compatible with the above machines Spo2 technology Cable lenth:- 3 meter approx Cable material:- TPU jacket material & latex free Spo2 Accuracy:- For 99 to 70% saturation ±2% accuracy Should be ISO/CE/BIS certified product 			

Warranty 1 year Comprehensive			
Item No.16 :- Multipara Monitor Make: COMEN Model: C-80 Sr. ➤ Specification For NIBP cuff (Adult):- <ul style="list-style-type: none"> • Compatible reusable NIBP cuff (Adult) with male connector pin. • Nylon,TPU material • Size: 25-35cm • Reusable with bladder • Hose length -1 ft approx. • Should be ISO/CE/BIS certified products only. 6 months comprehensive warranty			
Item No.17 :- Multipara Monitor Make: Comen Model: C-80 ➤ Specification For 5 lead ECG cable (adult) <ul style="list-style-type: none"> • Compatible reusable 5 lead ECG limb lead wires (ADULT) with trunk cable. • Should have grabber type limb wires. • should have detachable limb wires so that each wire can be replaced as and when required. • Lead should have color code as per IEC. • Should be ISO/CE/BIS certified product only. One year comprehensive warranty			
Item No.18:- Multipara Monitor(Bed side) Make:Philips Model: MX-450 ➤ Specification For Compatible reusable Clip Type/ Soft Rubber Adult SpO2 Probe. <ul style="list-style-type: none"> • The spo2 probe must be Compatible reusable spO2 probe soft rubber/clip typefor philips Fast Spo2Technology • cable material:- TPU jacket material & latex free. • Length: approx 3 mtr. • Spo2 Accuracy:- For 70 to 100 % saturation ±2% accuracy • Demonstration of the spo2 probe on patient in mandatory • Should be ISO/CE/BIS Certified products only. One year comprehensive warranty			
Item No.19:- Multipara Monitor(Bed side) Make:Philips Model: MX-450 ➤ Specification For Compatible 5 Lead ECG cable <ul style="list-style-type: none"> • Compatible reusable 5 ECG limb lead grabber type wires with trunk cable. • Should have grabber type limb wires. • Should have detachable limb wires(preferable) so that each wire can be replaced as and when 			

<p>required.</p> <ul style="list-style-type: none"> • All the lead wires must be encased in a lockable socket(if detachable) • Must be defibrillator shock proof. • Should be ISO/CE/BIS certified. <p>One year comprehensive warranty</p>			
<p>Item No.20 :- Multipara Monitor Make: Philips Model: : MX450</p> <p>➤ Specification For reusable NIBA Cuff (adult):-</p> <ul style="list-style-type: none"> • Compatible reusable NIBA Cuff (adult) with male connector pin. • Nylon, TPU material • Size:- 25-35cm • Reusable with bladder • Hose length-1 ft approx. • Should be ISO/CE/BIS Certified products only. <p>6 month comprehensive warranty.</p>			
<p>Item No.21 :- 12 Channel ECG Machine Make- BPL Medical technologies Pvt. Ltd. Model-9108 BPL cardiatic</p> <p>➤ Specification For Compatible 10 Lead ECG Patient cable</p> <ul style="list-style-type: none"> • Compatible reusable 10 Lead banana pin type ECG patient cable type ECG patient Cable. • The ECG cable should be labeled with standard electrode names for easy identification • should follow IEC (European standard) color coding standards for ECG cable • The 10 Lead ECG cable should be provided with junction for replacing /repairing lead wires easily & should not be with molded junction. • The bidders can check the unit onsite for better overview of the product. • Should supply with color coded limb electrodes-1 set =4 pcs. & chest bulb silicon type with mattle electrodes 1 set= 6pcs. • Should be ISO/CE/BIS Certified Product. <p>Warranty: One year Comprehensive warranty. If the cables are not repaired under warranty the same needs to be replaced in the warranty period starting from the date of installation.</p>			
<p>Item No.22 :- ICU ventilator Make: Schiller Healthcare India Pvt. Ltd. Model: Neumovent Graphnet TS</p> <p>➤ Specification For compatible Oxygen sensor:-</p> <ul style="list-style-type: none"> • Oxygen sensor must be compatible with the above model. • Measurement Range : 0 to 100% oxygen 			

<ul style="list-style-type: none"> • Accuracy: $\pm 1\%$ of full scale at constant temperature and pressure. • Response time : < 12 seconds • Operating humidity : 0 to 99% RH (non-condensing) • Operating temperature 0 degree to 40 degree • Storage temperature : 10 degree to 30 degree • Certificate : CE or ISO certified product <p>One year Comprehensive Warranty.</p>			
<p>Item No.23 :- HFO Ventilator Make: Stephan(Schiller india pvt. ltd.) Model: Sophie Sr.S0381400796</p> <p>➤ Specification For compatible Oxygen sensor:-</p> <ul style="list-style-type: none"> • Output Voltage :9-1105 mV • Measurement range:0 to 100% O2 • Output connector : 2pin AMP Plug • Accuracy Within +/- 1% of full scale at constant temperature and pressure • Response time <12 seconds • Operating humidity 0 to 99% RH (non-condensing) • Operating temperature 0 degree to 40 degree • Storage temperature 10 degree to 30 degree • Certificate CE or ISO certified product <p>One year Comprehensive Warranty</p>			

Medical Superintendent

SMIMER Hospital

Price-Bid

Sr. No.	Item Name	Qty.	Unit Rate(Rs.)
1.	Multipara Monitor Make: ComenModel: Comen C-50 ➤ Specification For Reusable NIBP cuff (Adult) <ul style="list-style-type: none">Compatible reusable NIBP Cuff (adult) with male connector pinNylon, TPU materialSize:-25-35cmReusable with bladderHose Length -1 ft Approx.Should be ISO/CE/BIS certified product. Warranty 6 months Comprehensive	1	
2.	Multipara Monitor Make: ComenModel: Comen C-50 ➤ Specification For 5 lead ECG cable (Adult) <ul style="list-style-type: none">Compatible reusable 5 lead ECG limb lead wires (adult) with trunk cableShould have grabber type limb wiresLeads should be color coded as per IECShould be ISO/CE/BIS certified products only.Bidders are requested to kindly visit the site before bidding to get a complete overview of the unit for ECG cable connections before bidding. One year Comprehensive warranty	1	
3.	Multipara Monitor Make: Comen Model: Comen C-50 ➤ Specification For Reusable NIBP cuff (Adult) <ul style="list-style-type: none">Compatible reusable NIBP Cuff (adult) with male connector pinNylon, TPU materialSize:-25-35cmReusable with bladderHose Length -1 ft Approx.Should be ISO/CE/BIS certified product. Warranty 6 months Comprehensive warranty	1	
4.	Multipara Monitor Make: ContecModel: CMS 8000 ➤ Specification For Reusable NIBP cuff (Adult) <ul style="list-style-type: none">Compatible reusable NIBP Cuff (adult) with male connector pinNylon, TPU materialSize:-25-35cmReusable with bladderHose Length -1 ft Approx.Should be ISO/CE/BIS certified product.Warranty 6 months Comprehensive warranty	1	
5	Multipara Monitor Make: ContecModel: CMS 8000 ➤ Specification For Compatible reusable Clip/Soft Rubber adult Spo2 probe <ul style="list-style-type: none">The spo2 probe must be compatible with the above machines Spo2 technologyCable lenth:- 3 meter approxCable material:- TPU jacket material & latex freeSpo2 Accuracy:- For 99 to 70% saturation ±2% accuracyShould be ISO/CE/BIS certified product. Warranty 1 year Comprehensive	1	
6.	Multiparamonitor Model:- CMS 8000 Make:- Contec ➤ Specification For 5 lead ECG cable (adult):- <ul style="list-style-type: none">Compatible reusable 5 lead ECG limb lead wires (ADULT) with trunk cable.Should have grabber type limb wires.Leads should be color coded as per IEC.Should be ISO/CE/BIS/ CDSCO certified products only.One year comprehensive warranty.The bidders are requested to kindly check the above said unit	1	

	onsite before quoting.		
7.	3 Channel ECG Machine Make :comen Model: CM300 ➤ Specification For Compitable 10 lead ECG Patient Cable <ul style="list-style-type: none"> • Compitable reusable 10 lead banana pin type ECG patient cable. • The ECG cable should be labeled with standard electrode names for easy identification. • Should follow IEC (European Standard) color coding standard for ECG cable. • The 10 lead ECG cable should be provide with junctionfor replacing/repairing lead wires easily & should not be with molded junction. • Should be ISO/CE/BIS/FDA certified. One Year Comprehensive warranty	1	
8.	3 channel ECG Machine Make: BPL Healthcare Pvt.LtdModel:- Cardiart 6208 view plus ➤ Specification For Compatible 10 lead ECG patient cable :- <ul style="list-style-type: none"> • Compatible reusable 10 lead banana pin type ECG patient cable. • The ECG cable should be labeled with standard electrode names for easy identification. • should follow IEC(European Standard) Color coding standards for ECG cable. • The 10 lead ECG cable should be provided with junction for replacing/repairing lead wires easily & should not be with molded junction. • Should be ISO/CE/BIS certified product Warranty 1 year Comprehensive	1	
9.	Multipara Monitor Make: DragerModel: Vista 120S ➤ Specification For NIBP cuff (Adult):- <ul style="list-style-type: none"> • Compatible reusable NIBP cuff (Adult) with male connector pin. • Nylon,TPU material • Size: 25-35cm • Reusable with bladder • Hose length -1 ft approx. • Should be ISO/CE/BIS certified products only. • 6 months comprehensive warranty. 	1	
10.	Multipara bedsidemonitor Make : DRAEGER Model: Vista 120S ➤ Specification For Compatible reusable Clip Type/ Soft Rubber Adult SpO2 Probe <ul style="list-style-type: none"> • Cable lenth:- 3 meter approx • Cable material:- TPU jacket material & latex free • Spo2 Accuracy:- For 99 to 70% saturation ±3% accuracy • Should be ISO/CE/BIS certified product. • Kindly visit the site for compelete over view of the machine/probe and spo2 technology • Warranty 1 year Comprehensive Warranty 	1	
11	Multipara bedsidemonitor Make : DRAEGER Model: Vista 120S ➤ Specification For Compatible 5 Lead ECG cable <ul style="list-style-type: none"> • Compatible reusable 5 lead grabber type limb wires without trunk cable. • Should have detachable limb wires(preferable) so that each wire can be replaced as and when required. • All the lead wires must be encased in a lockable socket(if detachable) • Must be defibrilator shock proof. Also should be suitable while cautery machine is used in operation theatre. • Should follow color coding as per IEC. • Should be ISO/CE/BIS/FDA certified. One year comprehensive warranty	1	
12	Defibrillator Make: Nasan Model: Sajeevani 1006 ➤ Specification For Compatible Reusable NIBP cuff (Adult):- <ul style="list-style-type: none"> • Compatible reusable NIBP cuff (Adult) with male connector pin. • Nylon,TPU material • Size: 25-35cm 	1	

	<ul style="list-style-type: none"> Reusable with bladder Hose length -1 ft approx. Should be ISO/CE/BIS certified products only. 6 months comprehensive warranty.		
13	Defibrillator Make: Nasan Model:-Sanjeevani 1006 <ul style="list-style-type: none"> ➤ Specification For Rechargeable battery Compatible rechargeable lithium ion battery for above mention unit 11.1 Volt,5200 mAH The bidders are requested to kindly check the above machine on site before bidding. Should be ISO/CE/BIS certified product. Warranty 1 year Comprehensive Warranty	1	
14	Defibrillator Make: Nasan Sr. no. 200653 <ul style="list-style-type: none"> ➤ Specification For Rechargeable Battery Compatible rechargeable Li-ion battery for above mentioned unit. Battery details :- 11.1v,5200 mAh. Should be ISO/CE/BIS or equivalent certified Products Only. The bidders are requested to kindly check the above machine on site before bidding. Warranty : 1 Year comprehensive 	1	
15	Multipara Monitor Make: COMEN Model: C-80 Sr. <ul style="list-style-type: none"> ➤ Specification For Compatible reusable Clip/Soft Rubber adult Spo2 probe The spo2 probe must be compatible with the above machines Spo2 technology Cable length:- 3 meter approx Cable material:- TPU jacket material & latex free Spo2 Accuracy:- For 99 to 70% saturation ±2% accuracy Should be ISO/CE/BIS certified product Warranty 1 year Comprehensive 	1	
16	Multipara Monitor Make: COMEN Model: C-80 Sr. <ul style="list-style-type: none"> ➤ Specification For NIBP cuff (Adult):- Compatible reusable NIBP cuff (Adult) with male connector pin. Nylon,TPU material Size: 25-35cm Reusable with bladder Hose length -1 ft approx. Should be ISO/CE/BIS certified products only. 6 months comprehensive warranty 	1	
17	Multipara Monitor Make: Comen Model: C-80 <ul style="list-style-type: none"> ➤ Specification For 5 lead ECG cable (adult) Compatible reusable 5 lead ECG limb lead wires (ADULT) with trunk cable. Should have grabber type limb wires. should have detachable limb wires so that each wire can be replaced as and when required. Lead should have color code as per IEC. Should be ISO/CE/BIS certified product only. One year comprehensive warranty	1	
18	Multipara Monitor(Bed side) Make:Philips Model: MX-450 <ul style="list-style-type: none"> ➤ Specification For Compatible reusable Clip Type/ Soft Rubber Adult SpO2 Probe. The spo2 probe must be Compatible reusable spO2 probe soft rubber/clip type for philips Fast Spo2 Technology cable material:- TPU jacket material & latex free. Length: approx 3 mtr. Spo2 Accuracy:- For 70 to 100 % saturation ±2% accuracy Demonstration of the spo2 probe on patient in mandatory Should be ISO/CE/BIS Certified products only. One year comprehensive warranty	1	

19	Multipara Monitor(Bed side) Make:Philips Model: MX-450 ➤ Specification For Compatible 5 Lead ECG cable <ul style="list-style-type: none"> Compatible reusable 5 ECG limb lead grabber type wires with trunk cable. Should have grabber type limb wires. Should have detachable limb wires(preferable) so that each wire can be replaced as and when required. All the lead wires must be encased in a lockable socket(if detachable) Must be defibrillator shock proof. Should be ISO/CE/BIS certified. One year comprehensive warranty	1	
20	Multipara Monitor Make: Philips Model: : MX450 ➤ Specification For reusable NIBA Cuff (adult):- <ul style="list-style-type: none"> Compatible reusable NIBA Cuff (adult) with male connector pin. Nylon, TPU material Size:- 25-35cm Reusable with bladder Hose length-1 ft approx. Should be ISO/CE/BIS Certified products only. 6 month comprehensive warranty.	1	
21	12 Channel ECG MachineMake- BPL Medical technologies Pvt. Ltd. Model-9108 BPL cardiatic ➤ Specification For Compatible 10 Lead ECG Patient cable <ul style="list-style-type: none"> Compatible reusable 10 Lead banana pin type ECG patient cable type ECG patient Cable. The ECG cable should be labeled with standard electrode names for easy identification should follow IEC (European standard) color coding standards for ECG cable The 10 Lead ECG cable should be provided with junction for replacing /repairing lead wires easily & should not be with molded junction. The bidders can check the unit onsite for better overview of the product. Should supply with color coded limb electrodes-1 set =4 pcs. & chest bulb silicon type with mattle electrodes 1 set= 6pcs. Should be ISO/CE/BIS Certified Product. Warranty: One year Comprehensive warranty. If the cables are not repaired under warranty the same needs to be replaced in the warranty period starting from the date of installation.	1	
22	ICU ventilator Make: Schiller Healthcare India Pvt. Ltd. Model: Neumovent Graphnet TS ➤ Specification For compatible Oxygen sensor:- <ul style="list-style-type: none"> Oxygen sensor must be compatible with the above model. Measurement Range : 0 to 100% oxygen Accuracy: ±1% of full scale at constant temperature and pressure. Response time : < 12 seconds Operating humidity : 0 to 99% RH (non-condensing) Operating temperature 0 degree to 40 degree Storage temperature : 10 degree to 30 degree Certificate : CE or ISO certified product One year Comprehensive Warranty.	1	
23	HFO Ventilator Make: Stephan(Schiller india pvt. Ltd.) Model: Sophie Sr.S0381400796 ➤ Specification For compatible Oxygen sensor:- <ul style="list-style-type: none"> Output Voltage :9-1105 mV Measurement range:0 to 100% O2 Output connector : 2pin AMP Plug Accuracy Within +/- 1% of full scale at constant temperature 	1	

	<div>and pressure</div> <ul style="list-style-type: none">• Response time <12 seconds• Operating humidity 0 to 99% RH (non-condensing)• Operating temperture 0 degree to 40 degree• Storage temperture 10 degree to 30 degree• Certificate CE or ISO certified product <div>One year Comprehensive Warranty</div>		
--	--	--	--

Note: Above mentioned Instrument should be physically checked in, Departments by the Quoting supplier prior to Quoting in SMIMER Hospital.

Medical Superintendent
SMIMER Hospital